

16.7-02 AN X-RAY HIGH TEMPERATURE CAMERA. By W. Adhart, H. Jagodzinski, S. Sueno and N. Tzafaras, Institut für Kristallographie und Mineralogie der Universität München, München, Federal Republic of Germany.

A heating camera has been developed for film and counter investigations of single crystals for a temperature range of 500 to 1900 K. Furthermore it is possible to apply various gas partial pressures. The oxygen partial pressure can be varied between 10^{-8} to 10^{-1} mbar in the temperature range of 500 to 1100 K. The camera is designed for the use of focussing monochromators. Various focussing conditions may be achieved by either turning around or translating along the incident beam (H. Jagodzinski, Acta Cryst. (1968) B24, 19). Since modified Weissenberg cameras are used, also Weissenberg oscillations and Noromosaic techniques (H. Jagodzinski and M. Korekawa, Geochim. Cosmochim. Acta Suppl. 4 (1975) Vol. 1, 955) can be performed. The camera consists mainly of a Be-tube (diameter = 57.5 mm, length = 110 mm, wall thickness = 0.5 mm), a tubular graphite heating element and a Pt - Pt, Rh thermocouple at the distance of 1 mm from the crystal. Depending on the precision of the gauging, a PID controller provides an absolute accuracy of 5 - 10 K, a stabilization of 0.2 K (at 400 K) and a reproducibility of 1 - 2 K (at 1550 K). The heating power is 250 W at $T = 1900$ K.

16.7-03 HIGH TEMPERATURE DIAMOND ANVIL CELL FOR SINGLE CRYSTAL X-RAY EXPERIMENTS. By S. Miyano and S. Sueno, Department of Geoscience, The University of Tsukuba, Ibaraki, 305, Japan.

A small diamond anvil cell with a small heater is useful for high-temperature and high-pressure single crystal X-ray experiments, especially for the in-situ observation of phase transition and recrystallization phenomena. We have designed a diamond cell 32mm in diameter, 12mm in thickness and 70g on weight, and tried to install several types of heater on it. A band type heater surrounding the outside of the cell showed the best results. The size of the cell with the heater is about 50mm in diameter, 30mm in thickness and 170g on weight. The maximum temperature so far we have obtained is 550°C (100V-100W). The whole system can be mounted on a normal goniometer-head and it is easy to take single crystal photographs on the standard type precession camera with a slight modification at the collimator and beam-stopper. One difficulty on the high-temperature and high-pressure work of the diamond cell is the pressure decrease at higher temperature, especially above 470°C . This may mainly be caused by the leaking of liquid pressure medium by the distortion or softening of the metal gasket. NaCl powder as pressure medium showed better results. This system is also very convenient for the in-situ observation of transition or recrystallization phenomena on the standard polarizing microscope with a small water cooling cover for protecting the objective lens from heat. The temperature of the specimen can be measured by thermo-couple attached by the diamond anvil. The pressure measurement at high temperature is difficult because the convenient ruby R_1 line shift method can not be applied above 400°C due to broadening of the R_1 -peak.

16.8-01 APPLICATION OF SYNTHESIZED HOLOGRAMS TO ELECTRON-MICROSCOPIC STUDIES OF CRYSTALS. By V.V. Aristov, G.A. Bashkina, A.I. Erko, L.S. Kohanchick, I.I. Khodos, Institute of Solid State Physics, Academy of Sciences of the USSR, Chernogolovka, Moscow district, 142432, USSR.

The paper investigates the applicability of holograms synthesized in respect to projections, obtained in EM to the measurements of various geometrical parameters of crystals. The measured parameters in TEM are the following: directions of dislocation lines, depth of defects occurrence and in SEM - relief of the surface. Relative accuracy of measurements may be as great as 1-0,1%.

The investigations performed show that the synthesized holograms (SH) look promising for getting information on the value of local electric and magnetic fields at the specimen surface. The measurements can be carried out over the pseudorelief of the surface arising on reconstruction of the SH.

16.8-02 LIGHT SCATTERING TOMOGRAPHY FOR DETECTION OF CRYSTALLINE DEFECTS.

By Kazuo Moriya and Tomoya Ogawa, Dept. of Physics, Gakushuin Univ., Mejiro, Tokyo, Japan.

Light scattering tomography was developed for the observations of clusters of hydrogen bonded OH caused by fluctuations in growth speed of synthetic quartz crystals (J. Cryst. Growth 44 (1987) 53), undecorated dislocation lines (Phil. Mag. 41 A (1980) 191) and plate-like defects in the crystals (Phil. Mag. A, in press) which act as phase objects.

Each tomograph was taken by a single scanning of a light pencil with about $20\ \mu\text{m}$ diam from a few mW He-Ne laser using a streak camera and ultra-microscope shown schematically in the figure.

Layer-by-layer tomographs obtained successively give a bird's eye view of defects in a crystal. The defects will be characterized by their image contrasts which are sharply dependent upon the scattering vector.

The results obtained in synthetic quartz & diamond crystals will be presented.

