

FA1-MS13-P09

Development of the Protein Microcrystal Handling Technique Using a Laser Tweezers. Takaaki Hikima^a, Tetsuya Shimizu^a, Masaki Yamamoto^a.
^a*RIKEN Spring-8 Center, Hyogo, Japan.*
 E-mail: hikima@spring8.or.jp

In synchrotron facilities, X-ray microbeam will be utilized for protein crystallography. It will be possible to collect diffraction data from a protein microcrystal with the size in the range from 1 μm to 10 μm . Until now a protein crystal is picked up manually from a crystallization droplet using a cryoloop and it mounted on a goniometer head. However it seems to be impossible to manipulate the protein microcrystals are very small and fragile against a shock. So we are developing an automatic microcrystal pick-up system. In the system, we applied laser tweezers to manipulate the fragile protein microcrystal. It was reported that the laser tweezers at the near-infrared region traps and manipulates a cell without critical photodamage. We applied a laser tweezers with two single-mode lensed fiber probes. The system could manipulate the protein microcrystal with lower emission power than the laser tweezers based on conventional condensing lens. The lensed fiber probe was small, which had an advantage to manipulate a microcrystal in the crystallization droplets at various crystallization plates. The laser tweezers succeeded in trapping protein crystals smaller than 25 μm and levitating it onto the cryoloop. X-ray measurement of the manipulated protein microcrystals at SPring-8 BL41XU indicated that laser trap with 1064 nm wavelength hardly affected the result of X-ray structural analysis.

Keywords: crystallography of protein small molecules; laser technology; microcrystals

FA1-MS13-P10

Crystal Structures of Metal Ion Containing Adenylate Kinase from *Desulfovibrio Gigas*. A. Mukhopadhyay^a, A.V. Kladova^a, J. Trincão^a, S.A. Bursakov^{a,b}, I. Moura^a, J.J.G. Moura^a, M.J. Romão^a.
^a*REQUIMTE, Departamento de Quimica, Centro de Quimica Fina e Biotecnologia, Faculdade de Ciências e Tecnologia, Universidade Nova de Lisboa, 2829-516 Caparica, Portugal.* ^b*Departamento de Protección Ambiental, Estación Experimental del Zaidín, CSIC, C/Profesor Albareda 1, 18008 Granada, Spain.*
 E-mail: abhik@dq.fct.unl.pt

Adenylate kinases (AK) from Gram-negative bacteria are usually devoid of metal ions in their LID domain. Crystal structures of substrate free AK from *Desulfovibrio gigas* (AK_{gig}) with three different metal ions (Zn²⁺, Co²⁺ and Fe²⁺) bound in its LID domain have been determined by X-ray crystallography to resolutions 2.1, 2.0 and 3.0 Å, respectively. The zinc and iron forms of the enzyme crystallized in the space group *I*222, whereas the cobalt form crystals were *C*2. The presence of the metals was confirmed

by calculation of anomalous difference maps and by X-ray fluorescence scans. The work presented herein is the first report of a structure of a metal containing Gram-negative. Native Co/Zn enzyme was crystallized, but only zinc was detected in its LID domain. Co- and Fe- AK forms were obtained by overexpressing the protein in *E. coli* grown in minimal medium supplemented with the appropriate MCl₂ (where M is Co²⁺, Zn²⁺, Fe²⁺). Zn- and Fe- forms of AK crystallized as monomers in the asymmetric unit, whereas the Co-AK crystallized as a back to back dimer. However, all three crystal structures are very similar to each other with the same LID domain topology with the only change being the presence of different metal atoms in it. Absence of fully occupied metal sites in all three structures may indicate the weak binding of the metal atom to the enzyme. In the absence of any substrate, the LID domain of all holo forms of AK was present in a fully open conformation state. The normal mode analysis was performed and fluctuation of the LID domain along the catalytic pathway was predicted.

[1] Gavel O. Yu., Bursakov S.A., Rocco G. D., Trincão J., Pickering I. J., George G. N., Calvete J. J., Shnyrov V. L., Brondino C. D., Pereira A. S., Lampreia J., Tavares P., Moura J. J.G., Moura I., 2008, *Journal of Inorganic Biochemistry*, 102, 1380–1395.

Keywords: metalloenzyme X-ray crystallography; MAD phasing; gram negative bacteria

FA1-MS13-P11

Proton Transfer Mechanisms in Multi-copper Oxidases: Studies in CotA-laccase. Catarina S. Silva^a, Zhenjia Chen^a, Paulo Durão^a, Peter F. Lindley^a, Lígia O. Martins^a, Cláudio M. Soares^a, Isabel Bento^a.
^a*Instituto de Tecnologia Química e Biológica, Universidade Nova de Lisboa, Av. da República, EAN, 2781-901 Oeiras, Portugal.*
 E-mail: cisilva@itqb.unl.pt

Multi-copper oxidases (MCOs) constitute a family of enzymes responsible for coupling substrate oxidation with the reduction of dioxygen to water. Widely distributed in nature, MCOs include ascorbate oxidase, ceruloplasmin and laccases, the latter ones being the simplest representative members of this family of enzymes. As observed amongst its members, these enzymes contain at least two different copper centres as their minimal functional unit: a mononuclear type 1 blue copper centre (T1) and a trinuclear cluster comprising two type 3 and one type 2 copper atoms (T2/T3 site). The substrate oxidation occurs at the mononuclear centre, shuttling electrons to the trinuclear centre where dioxygen binding and reduction occurs along with the production of water molecules.

Using CotA-laccase as a model system, we have proposed a putative mechanism of oxygen reduction for this type of enzyme [1]. However, many questions relating to its catalytic mechanism remain to be addressed.

In nearly all multi-copper oxidases it is observable the existence of a carboxylated group conserved in the neighbouring of the trinuclear cluster, suggested to be involved in protonation events [2], [3]. Our CE/MC

calculations on the different states of the mechanism show differences in the degree of protonation of this residue, showing that it may provide protons necessary for the initial steps of oxygen reduction to water. To corroborate these findings we have studied the role of glutamate 498 (E498), through crystal structure determination of three different mutant enzymes in which this residue was replaced by site-directed mutagenesis, in an attempt to further understand several structural and functional aspects of the enzyme mechanism. Being located within hydrogen bonding distance of a water molecule in the dioxygen entrance channel and directly interacting with the dioxygen moiety that binds between the two type 3 copper atoms, not only we confirm that E498 acts as an important proton donor during the catalytic mechanism, but we also take a step further demonstrating the role of this residue in the stabilization of the dioxygen reduction site as a whole.

[1] Bento I., Carrondo M.A., Lindley P.F., *J. Biol. Inorg. Chem.*, **2006**, 11, 539. [2] Kataoka K., Sugiyama R., Hirota S., Inoue M., Urata K., Minagawa Y., Seo D., Sakurai T., *J. Biol. Chem.*, **2009**, ahead of printing. [3] Solomon E.I., Augustine A.J., Yoon J., *Dalton Trans.*, **2008**, 30, 3921.

Keywords: CotA laccase; site-directed mutagenesis; biomolecular modelling

FA1-MS13-P12

The Phase 1 MX Beamlines at Diamond Light Source: Experience from Operation and Commissioning. Ralf Flaig^a, Macromolecular Crystallography Group^a. ^a*Diamond Light Source, Harwell Science and Innovation Campus, Chilton, Didcot, Oxfordshire, OX11 0DE, UK.*
E-mail: ralf.flraig@diamond.ac.uk

Diamond Light Source [1] is the new UK third generation synchrotron located south of Oxford. In January 2007 Diamond welcomed first users.

In Phase 1 seven beamlines are funded which includes three beamlines for macromolecular crystallography (MX) [2]. These are currently running a user programme for both academic and industrial users. Commissioning time is also scheduled with the aim for optimisation of operation. The beamlines are similar in design and take radiation from an in-vacuum undulator. A double crystal monochromator and a Kirkpatrick-Baez mirror arrangement are the main optical components. All beamlines are fully tuneable and are equipped with automatic sample changers.

Experience and results from operation and commissioning of the MX beamlines will be presented. This will include discussion of the beam properties, status and performance of the optical components and diagnostics in the optics hutch as well as results from commissioning of the equipment in the experimental end station. The software environment and results from data collections will also be discussed. Latest developments and a future outlook will be presented.

[1] <http://www.diamond.ac.uk>. [2] <http://www.diamond.ac.uk/MX>

Keywords: diamond light source; macromolecular crystallography; beamlines