

ANKAphase: software for single-distance phase retrieval from inline X-ray phase-contrast radiographs. Erratum**T. Weitkamp,^{a*} D. Haas,^{b,c} D. Wegrzynek^{d,‡} and A. Rack^{e*}**

^aSynchrotron Soleil, Gif-sur-Yvette, France, ^bInstitute for Synchrotron Radiation/ ANKA Light Source, KIT, Karlsruhe, Germany, ^cHochschule Darmstadt, Darmstadt, Germany, ^dInternational Atomic Energy Agency, Seibersdorf, Austria, and ^eEuropean Synchrotron Radiation Facility, Grenoble, France.
E-mail: weitkamp@synchrotron-soleil.fr, arack@snaifu.de

An equation in the paper by Weitkamp *et al.* [(2011), *J. Synchrotron Rad.* **18**, 617–629] is corrected.

There was a mistake in equation (6) in the paper by Weitkamp *et al.* (2011). The equation should read

$$\varphi(x, y) = \frac{\delta}{2\beta} \ln \left(\mathcal{F}^{-1} \left\{ \frac{\mathcal{F}[I(x, y)/I_0(x, y)]}{1 + [\lambda z \delta / (4\pi\beta)](u^2 + v^2)} \right\} \right). \quad (6)$$

References

Weitkamp, T., Haas, D., Wegrzynek, D. & Rack, A. (2011). *J. Synchrotron Rad.* **18**, 617–629.

[‡] Present address: University of Science and Technology, Kraków, Poland.