305-0047, Japan, E-mail:sakurai@yuhgiri.nims.go.jp

The present talk describes novel powerful imaging for X-ray fluorescence (XRF) and X-ray diffraction (XRD). So far, the scanning-type imaging has been widely used in those techniques. Though recent progress in high-spatial-resolution imaging using synchrotrons is significant, there have been a clear limit; because of the step-scan, the imaging requires a long measuring time. In many scientific applications, X-ray imaging that are much more rapid, e.g., capable of high-speed resolution have been demanded. It is possible to do X-ray imaging without performing any scans. Here, the method uses quite a wide beam, which illuminates the whole sample surface in a low-angle-incidence arrangement (0.5~3 deg). The detector used is a CCD camera working at 30 fr./sec, equipped with a collimator inside, and the distance between the sample surface and the detector is set extremely close, in order to enhance both spatial resolution and efficiency. Note that the imaging is done with one shot. In the case of XRF imaging, distinguishing elements are required and, therefore, most of the experiments were performed with monochromatic or quasi-monochromatic X-rays. The procedure for XRD imaging uses a combination of exposure and incident X-ray energy scan (or just tuning). Since the present experiment employs a fixed small-angle incidence and also a fixed diffraction angle of around 90 deg, the diffraction plane here is inclined at about 45 deg from the surface of the specimen. By scanning the energy of the incident X-rays, one obtains a diffraction peak, which corresponds to the lattice spacing. Further instrumental details and many applications will be presented. References

[1] K.Sakurai, Spectrochimica Acta B54, 1497 (1999)

[2] K.Sakurai and H.Eba, Anal. Chem. 75, 355 (2003)

Keywords: X-ray imaging, rapid X-ray measurement system, CCD detectors

### P01.07.43

Acta Cryst. (2008). A64, C184

### Low noise multichannel ASIC for readout of SSD used in diffraction for powder and multilayer samples

Piotr Maj, Pawel Grybos, Robert Szczygiel, Maciej Kachel, Piotr Kmon, Tomasz Stobiecki

AGH University of Science and Technology, Measurement and Instrumentation, Tokarskiego 4/215, Krakow, Krakow, 30-065, Poland, E-mail:piotr.maj@agh.edu.pl

We present 64 channel integrated circuit designed in CMOS 0.35  $\mu$ m technology for readout of silicon strip detector used in X-ray diffractometry applications. This integrated circuit called DEDIX (Dual Energy Digital Imaging of X-rays) connected to silicon strip detector works in a single photon counting mode and is able to select photons from a given energy window. The main parameters of this ASIC are low noise performance (110 el. rms for Cdet=1pF), high count rate capability (up to 1 Mcps/channel) and very good channel to channel uniformity (effective threshold spread below 7 el. rms). Using the ASICs we have built several 128-channel modules with silicon strip detectors of different strip length (1 or 2 cm) and different strip pitch (50 µm, 75 µm and 100 µm). We have used these modules in diffractometry applications as a replacement of a proportional counter. The measurements were speeded up by over 100 times when using our module. These modules have been tested using different samples, where we compare: results obtained with and without photon selection in a given energy window, the influence of silicon detector strip pitch, measurement, when the module fixed on the diffractometer arm is moving in step or continuously scan mode. We also performed some measurements of multilayer structures,

and due to low noise, good uniformity of channels and high rate capability, the measurement time have been significantly reduced without loosing the diffractogram quality.

Keywords: silicon strip detector, integrated circuit, DEDIX

## P01.08.44

Acta Cryst. (2008). A64, C184

# A new biological neutron diffractometer (iBIX) in J-PARC

<u>Ichiro Tanaka</u><sup>1</sup>, Katsuhiro Kusaka<sup>1</sup>, Katsuaki Tomoyori<sup>1</sup>, Nobuo Niimura<sup>1</sup>, Takashi Ohhara<sup>2</sup>, Kazuo Kurihara<sup>2</sup>, Takaaki Hosoya<sup>2</sup>, Tomoji Ozeki<sup>3</sup>

<sup>1</sup>Ibaraki University, Faculty of Engineering, 4-12-1, Hitachi, Ibaraki, 316-8511, Japan, <sup>2</sup>Japan Atomic Energy Agency (JAEA), 2-4, Shirakata-Shirane, Tokai, Ibaraki, 319-1195, Japan, <sup>3</sup>Tokyo Institute of Technology, 2-12-1, Ookayama, Meguro, Tokyo, 152-8551, Japan, E-mail : i.tanaka@mx.ibaraki.ac.jp

Since 2004, Ibaraki Prefectural Government in Japan has started to construct a TOF neutron diffractometer for biological macromolecules for industrial use at J-PARC, near JRR-3 in JAEA. From December in 2008, Ibaraki University will operate this machine with a support of Ibaraki Prefecture. The diffractometer is designed to cover sample crystals which have their cell edges up to around 150 Å. It is expected to measure more than 100 samples per year if they have 2mm<sup>3</sup> in crystal volume. The efficiency is more than 100 times larger than the present high performance diffractometers, BIX-4 in JRR-3 reactor in JAEA. To realize this performance, a coupled moderator (intense neutrons, but broad pulse in time resolution) was selected. In addition, two important and key items should be developed; a new detector with high spatial resolution (less than 1mm) and a special software to de-convolute overlapped Bragg reflections in data reduction. The detector uses ZnS:Ag/6LiF scintillator with wavelength-shift-fiber (WLSF) system. The software has been designed using a complicated kind of profile-fitting method. The current status of the construction and developments will be reported.

Keywords: J-PARC, protein crystallography, TOF pulsed neutron diffractometer

## P01.08.45

Acta Cryst. (2008). A64, C184-185

### Optimization of design parameters of IBARAKI Biological Crystal Diffractometer (iBIX) in J-PARC

Katsuhiro Kusaka<sup>1</sup>, Takashi Ohhara<sup>2</sup>, kazuo Kurihara<sup>2</sup>, Ichiro Tanaka<sup>1</sup>, Takaaki Hosoya<sup>1</sup>, Katsuaki Tomoyori<sup>1</sup>, Tomoji Ozeki<sup>3</sup>, Nobuo Niimura<sup>1</sup>

<sup>1</sup>Ibaraki University, 4-12-1 Nakanarusawa, Hitachi, Ibaraki, 3168511, Japan, <sup>2</sup>Japan Atomic Energy Agency, 2-4 Shirakata-Shirane, Tokai, Ibaraki, 319-119, Japan, <sup>3</sup>Tokyo Institute of Technology, 2-12-1 Ookayama, Meguro-ku, Tokyo 152-8550, Japan, E-mail : kusakats@mx.ibaraki.ac.jp

IBARAKI prefecture in Japan has started to construct the TOF neutron biological diffractometer (IBARAKI Biological Crystal Diffractometer: iBIX) for industrial use at BL03, MLF in J-PARC. The diffractometer is designed to cover the samples have their cell edges up to around 150 angstrom and to achieve the efficiency which is more than 100 times larger than the present high performance diffractometer, BIX-3 (JRR-3, JAEA, Japan) in order to pioneer