

## MS20-O4

### Synchrotron single-crystal diffraction for chemistry and materials: new developments

William Clegg<sup>1</sup>

1. Chemistry, School of Natural and Environmental Sciences,  
Newcastle University, Newcastle upon Tyne, United Kingdom

**email:** [bill.clegg@ncl.ac.uk](mailto:bill.clegg@ncl.ac.uk)

Prior to the 1990s there were no dedicated single-crystal diffraction beamlines for use by chemists and materials scientists; limited access was available to multi-purpose beamlines used also, and in most cases predominantly, for powder diffraction and/or spectroscopic techniques, along with some minor use of macromolecular crystallography beamlines. The design, construction, and commissioning of Station 9.8 at Daresbury Laboratory's Synchrotron Radiation Source (SRS)<sup>1</sup> thus began what is now a rich though still brief 20-year history of bringing the advantages of synchrotron radiation to the community of chemical crystallographers familiar with this key technique in a standard local laboratory setting. The advantages derive from the special properties of synchrotron radiation, principally its high intensity, excellent collimation, and available range of wavelengths.

While the UK led the way, initially at SRS with Station 9.8 and its later more modestly configured Station 16.2SMX, the rapid uptake and exploitation of the new opportunities prompted similar development of dedicated and optimised single-crystal diffraction beamlines elsewhere, notably at ALS, together with enhancement of single-crystal diffraction facilities on shared beamlines such as those at ESRF, APS, and Soleil. With the closure of SRS, a new UK beamline (I19) was designed and constructed at Diamond Light Source (DLS), beginning operation in 2008. This has recently undergone a major upgrade.<sup>2</sup>

These beamlines have been used for structural studies across all branches of organic, inorganic, materials, and pharmaceutical chemistry, and have played a vital role in the investigation of large-scale supramolecular assemblies and extended framework structures, with applications in gas and energy storage among other important modern research areas.

Our own use has included the elucidation of structures displayed on journal front covers, some unusual structural features such as highly strained metallated aromatic molecules, and a natural product with exceptional antibiotic properties. Recent achievements include the successful development of remote access operation of DLS beamline I19,<sup>3</sup> which should now become the standard form of usage of the beamline, bringing many advantages. Examples will be shown, along with a summary of the history and new aspects.

#### References:

- 
- [1]. Cernik, R. J., Clegg, W., Catlow, C. R. A., Bushnell-Wye, G., Flaherty, J. V., Greaves, G. N., Burrows, I., Taylor, D. J., Teat, S. J. & Hamichi, M. (1997). *J. Synchrotron Rad.*, 4, 279–286.
- [2]. Allan, D. R., Nowell, H., Barnett, S. A., Warren, M. R., Wilcox, A., Christensen, J., Saunders, L. K., Peach, A., Hooper, M. T., Zaja, L., Patel, S., Cahill, L., Marshall, R., Trimmell, S., Foster, A. J., Bates, T., Lay, S., Williams, M. A., Hathaway, P. V., Winter, G., Gerstel, M. & Wooley, R.W. (2017). *Crystals*, 7, 336.
- [3]. Johnson, N.T., Waddell, P. G., Clegg, W. & Probert, M. R. (2017). *Crystals*, 7, 360.
- 

**Keywords:** [Synchrotron](#), [remote access](#), [data collection](#)