STRUCTURAL BIOLOGY
ISSN 2059-7983

QwenCryoMarker: a universal post-processing framework for contamination-aware particle cleaning. Corrigendum

Yunhai Sun,[a,b] Jiahao Zhao,[c,d] Nan Xu,[c,d] Lei Wang,[c,d] Wei Ding[a,b]* and Ming Li[a,b]*

[a]Laboratory of Soft Matter Physics, Institute of Physics, Chinese Academy of Sciences, Beijing 100190, People's Republic of China, [b]School of Physical Sciences, University of Chinese Academy of Sciences, Beijing 100049, People's Republic of China, [c]Institute of Automation, Chinese Academy of Sciences, Beijing 100190, People's Republic of China, and [d]ScienceOne Team, Beijing 100190, People's Republic of China. *Correspondence e-mail: dingwei@iphy.ac.cn, mingli@iphy.ac.cn

Received 6 November 2026
Accepted 6 November 2026

Edited by J. Wang, Tsinghua University, People's Republic of China

Keywords: cryo-EM; contamination marking; vision-language model; post-processing; *QwenCryoMarker*.

Two corrections are made to Table 6 in the article by Sun *et al.* (2026).

Two values in Table 6 in the article by Sun *et al.* (2026) were given incorrectly. The correct values for EMPIAR-10017/*Topaz* are 3.97 Å before cleaning and 3.88 Å after cleaning. The corrected Table 6 is presented below.

Table 6
3D refinement resolution (Å) before and after applying *QwenCryoMarker* for all dataset–picker pairs.

	CryoTransformer		Blob detection		Template matching		*Topaz*	
EMPIAR ID	Before	After	Before	After	Before	After	Before	After
EMPIAR-10017	4.36	4.10	4.01	4.01	3.92	3.97	3.97	3.88
EMPIAR-10081	4.33	4.22	3.58	3.61	3.56	3.58	3.62	3.62
EMPIAR-10093	6.22	6.20	4.30	4.42	3.74	3.71	5.21	5.18
EMPIAR-10345	5.42	5.00	6.35	6.26	7.00	6.93	6.93	6.43
EMPIAR-10532	3.32	3.24	7.73	6.57	3.62	3.32	3.77	3.75

References

Sun, Y., Zhao, J., Xu, N., Wang, L., Ding, W. & Li, M. (2026). *Acta Cryst.* D**82**, 1093–1110.

